

DATASHEET

3rd sense electric surgery table - N° TA700 & TA703 - range 110/111/115


This surgery table allows to operate different kinds of animals. It is made of 304L stainless steel, ensuring it a long life durability, perfect for a sterile environment, because easy to clean and sanitize and corrosion resistant.

Adjusting the animal's position is made easy thanks to the height adjustment with a simple foot control. The easy and smooth Tredelenburg adjustment, also electric thanks to its foot switch, enables a direct access to the surgical field.

The tray can also tilt in its width thanks to the "3rd sense" tilt, which gives even more ease during the procedures.

The two tie-down bars on each side allow both to strategically hold the animal during the process, while ensuring its full safety, as well as to be able to position essential accessories for surgeries.

To ensure an optimal comfort, you can, in addition to the flat tray, choose a tray with 1 or 2 holes or also choose a X-ray one. The table can also be fitted with castors for easy movements. (+ option when the order is placed - output socket under the pyramidal base)



- Electric actuator for smooth and silent up-and-down operations
- Electric foot switch for height adjustment
- Electric Tredelenburg adjustments (2nd sense)
- Electric foot switch for the width tilt (3rd sense)
- Tilt angle in lenght = 30° forward & 60° backward
- Tilt angle in width = 45° on each side
- Stainless steel top with height adjustment from 58 to 115 cm
- Tie-down bar on each side + tie-down hooks
- "Z" frame design
- Entirely in 304 L stainless steel
- All electrics are certified for EU standards
- 5 years warranty for the frame
- 2 years warranty for electric parts
- *Removable castors by manual foot switch diameter of 75 mm
- ** X-ray option blue polymer coating 16 mm thick

References:

TA 700110 - 3rd sense electric surgery table - flat tray + 4 castors*

 $\it TA~700111$ - 3rd sense electric surgery table - tray with 1 hole + 4 castors*

TA 700115 - 3rd sense electric surgery table - tray with 2 holes + 4 castors*

TA 703110 - 3rd sense electric surgery table - X-ray flat tray + 4 castors*

TA 703111 - 3rd sense electric surgery table - X-ray tray** with 1 hole + 4 castors*

TA 703115 - 3rd sense electric surgery table - X-ray tray** with 2 holes + 4 castors*

Dimensions:

• Table top: 140 x 53 cm

• Minimal height: 58 cm

• Maximal height: 115 cm

Maximum load: 100 KG

Custom sizes can be made to order







USER MANUAL

3rd sense electric surgery table - N° TA700 & TA703 - range 110/111/115

Please read carefully the user manual before operating the device

First of all, don't forget to activate your warranty directly on our website www.vetinox.fr - tab Product Warrantv.

Then, let yourself be guided to register your device.

• Set-up - Activation

- Carefully unpack the device
- Remove the foam wedges
- Connect the table to a power socket protected by a residual-current circuit breaker
- Warning: do not use the Tredelenburg adjustements in the low position
- To lift the table, simply press the foot switch with an arrow up
- To lower the table, simply press the foot switch with an arrow down
- To tilt the table lengthwise, simply press the relevant foot switch
- To tilt the table in its width, simply press the relevant foot switch
- To activate the castors under the base, press the manual foot switch on the right side, on the rear width of the table

• Maintenance - in case of failure

- A warranty of 5 years is offered to cover mechanical defects or corrosion as well as a warranty of 2 years for electric parts.
- This is a geared motor actuator power 220 Volts 400 Watts.
- Current-power = 220-240 volts.

Your table does not rise or lower anymore, or no longer tilts:

- Check if the power cable is properly connected or if it has not been severed
- Check your circuit breaker on your power switchboard
- Check the fuse within the table connexion box
- If in spite of all these checks, your table will not operate, please contact us, we will be pleased to take care of your problem

• Contraindications

- Do not immerse the table or wash it with a high pressure spray shock hazards
- Do not exceed the maximum authorized weight
- Maximum allowable weight: 100 KG
- Connect the device to a secure electrical network
- Do not change the existing frame or the electrical system and in general do not make any modification
- Install the device in a suitable clean place

• Cleaning and precautions

The stainless steel used is type 304 L. This stainless steel has a very high resistance to corrosion. As stainless steel is a solid material and not a surface coating, the appearance of stains or traces may not be linked to rust in any case.

You can clean and restore the brushed appearance with a synthetic scouring pad, like green Scotch-Brit, by scrubbing lightly with the grain. Do not use steel wool. Ban all products in cream formulation which might clog or bleach and chlorine - based cleaners. It is better to favor glass products or quaternary ammonium.

If you have any doubt, please send us a photo of the cleaning product composition to confirm use.



